



EFFICACY OF THE TADIRAN AIR CARE O2-1 AGAINST AEROSOLIZED SARS-COV-2

PROJECT: TADIRAN AIR CARE O2-1 - SARS-COV-2

PRODUCT: AIR CARE O2-1

CAP LIC NO: 8860298

CLIA LIC NO: 05D0955926

STATE ID: CLF 00324630

CHALLENGE ORGANISM(S):

SARS-COV-2 USA-CA1/2020

Medical Director

Dana Yee, M.D.

Testing Facility

Innovative Bioanalysis, Inc.

3188 Airway Ave Suite D

Costa Mesa, CA 92626

www.InnovativeBioanalysis.com

Email: info@innovativebioanalysis.com

Laboratory Project Number

1190



Table of Contents

EFFICACY OF THE TADIRAN AIR CARE O2-1 AGAINST AEROSOLIZED SARS-COV-2	1
Efficacy Study Summary.....	3
Study Report	4
Study Title:	4
Sponsor:	4
Test Facility:	4
Device Testing:	4
Study Report Date: 11/15/2021.....	4
Experimental Start Date: 10/14/2021.....	4
Experimental End Date: 10/21/2021	4
Study Completion Date: 11/14/2021	4
Study Objective:	4
Test Method:.....	4
Test System Strains:	4
Study Materials and Equipment:	5
Test Method:.....	7
Control Protocol.....	9
Study Results.....	9
Conclusion.....	10
Disclaimer.....	11



Efficacy Study Summary

Study Title	EFFICACY OF THE TADIRAN AIR CARE O2-1 AGAINST AEROSOLIZED SARS-COV-2
Laboratory Project #	1190
Guideline:	Modified ISO standards as no test standards exist for this testing.
Testing Facility	Innovative Bioanalysis, Inc.
GLP Compliance	All internal SOPs and processes follow GCLP guidelines and recommendations.
Test Substance	SARS-CoV-2 USA-CA1/2020
Description	The Tadiran Air Care O2-1 was designed to emit H ₂ O ₂ molecules to decrease concentrations of pathogens in the air while operational. The Air Care O2-1 is built to the technologies of Air Care O2. This in vitro study sought to evaluate the device's ability to reduce aerosolized SARS-CoV-2 within a 1-meter cubed test box.
Test Conditions	The test was conducted in a sealed 1-m ³ test box located inside a BSL-3 chamber. The temperature during testing was approximately 76 ±4°F (24 ±2°C), with a relative humidity of 31 ±4%. The nebulizer was filled with the 7.02 x 10 ⁶ Median Tissue Culture Infectious Dose per mL (TCID ₅₀ /mL) in viral media and nebulized at a constant rate into the testing chamber. Air samples were collected 10 minutes after nebulization stopped (T-0) 30, 60, 90, 120, and 150 minutes after running the device.
Test Results	At 30 minutes, the Air Care O2 unit reduced the starting concentration from 7.02 x 10 ⁶ TCID ₅₀ /mL to 8.46 x 10 ⁵ TCID ₅₀ /mL. As time elapsed, the device consistently reduced collectible aerosolized SARS-CoV-2 as seen by the amount collected at 60 minutes (9.60 x 10 ³ TCID ₅₀ /mL) and 90 minutes (1.20 x 10 ² TCID ₅₀ /mL).
Control Results	With the Air Care O2-1 unit not operating in the test environment, there was a 29.28% reduction at 60 minutes and a 69.74% reduction at 150 minutes. The results displayed a natural viability loss in the chamber and were used as a comparative baseline to calculate net viral reduction.
Conclusion	Overall, the Air Care O2-1 system as it was set up demonstrated observable ability to reduce aerosolized SARS-CoV-2 compared to natural loss rates. The device achieved the following reductions: 87.947% at 30 minutes, 99.863% at 60 minutes, and 99.998% at 90, 120 and 150 minutes of operation.



Study Report

Study Title: EFFICACY OF THE TADIRAN AIR CARE O2-1 AGAINST AEROSOLIZED SARS-COV-2

Sponsor: Tadiran Consumer & Technology Products Ltd.

Test Facility: Innovative Bioanalysis, Inc. 3188 Airway Ave Suite D, Costa Mesa, CA 92626

Device Testing: Air Care O2-1

Study Report Date: 11/15/2021

Experimental Start Date: 10/14/2021

Experimental End Date: 10/21/2021

Study Completion Date: 11/14/2021

Study Objective:

Tadiran provided the Air Care O2-1 for testing purposes to determine efficacy against viral pathogens. This study evaluated the effectiveness of the Air Care O2-1 at reducing the viral strain referred to as SARS-CoV-2 USA-CA1/2020 within the air.

Test Method:

Bioaerosol Generation:

The nebulizer was filled with a 7.02×10^6 Median Tissue Culture Infectious Dose (TCID₅₀) per mL viral media of SARS-CoV-2 and nebulized at a 1 mL/min flow rate with untreated local atmospheric air. The average particle size was approximately 0.8 μm . The nebulizer's remaining viral stock volume was weighed to confirm that roughly the same amount was nebulized during each run. Bioaerosol procedures for the controls and viral challenges were performed in the same manner with corresponding time points and collection rates.

Bioaerosol Sampling:

This study used three probes for air sampling, each connected to a calibrated Gilian 10i vacuum device. Before use, the devices were inspected for functionality. The air sampler operated in conjunction with a removable sealed cassette and manually removed after each time point. Cassettes had a delicate internal filtration disc to collect viral samples, which was moistened with viral suspension media to aid in the collection. The filtration disc from Zefon International, Lot# 26338, was used.

Test System Strains: SARS-CoV-2 USA-CA1/2020

The following reagent was deposited by the Centers for Disease Control and Prevention and obtained through BEI Resources, NIAID, NIH: SARS-Related Coronavirus 2, Isolate USA-CA1/2020, NR-52382.

INNOVATIVE BIOANALYSIS

creating solutions | getting results

Study Materials and Equipment:

Equipment Overview:

The equipment arrived at the laboratory pre-packaged from the manufacturer and was inspected for damage upon arrival. The system came with a power supply control box, preset fan, and internal components, all assembled and installed before arrival at the laboratory. The device was powered on to confirm functionality and safety before testing using the product setup procedure provided by the manufacturer. RKI air monitoring systems continuously sampled the air for O₃, H₂O₂, N₂O production to ensure safe working conditions for staff. No alarms for elevated O₃ were activated during testing. Air sampling for H₂O₂ was not designed for sub 50ppb measurements and is only a guideline.

MANUFACTURER: Tadiran Consumer & Technology Products Ltd.

MODEL: Air Care O2-1

SIZE: 4.98" x 4.05" x 2.95"

MAKE: N/A

PN: 51619202400



Testing Layout:

Testing was conducted inside a sealed 1-m³ test box located inside a BSL-3 room that followed BSL-3 standards. The chamber remained closed to prevent any air from entering and leaving the room during testing. The device with an attached preset fan was placed in the center of the test box on top of a support structure. A nebulizing port connected to a programmable compressor system was located behind and above the testing device, as seen in Figure 1. The test box was equipped with three probes positioned along the centerline of the top of the test chamber and protruded 10" down into the testing chamber. Before testing, the box and room were pressure tested and visually inspected for leaks. Also, all equipment required for testing underwent a function test to confirm proper working conditions.

INNOVATIVE BIOANALYSIS

creating solutions | getting results




Figure 1. Testing layout for control and experimental trials.



Test Method:

Exposure Conditions:

1. The temperature during all test runs was approximately $76 \pm 4^{\circ}\text{F}$ ($24 \pm 2^{\circ}\text{C}$) with a relative humidity of $31 \pm 4\%$.
2. Samples were collected after nebulization stopped for 10-minutes (T-0) at 30, 60, 90, 120, and 150 minutes.

Experimental Procedure:

1. Before the initial control test and following each trial, the testing area was decontaminated and prepped per internal procedures.
2. 1 mL of a 7.02×10^6 TCID₅₀/mL of SARS-CoV-2 viral media was nebulized via a dissemination port into the room.
3. After nebulization, the Air Care O2-1 unit was turned on via remote control.
4. At each predetermined time point, the device was turned off for sample collection.
5. Air sampling collections were set to 10-minute continuous draws at the point of sampling.
6. Sample cassettes were manually removed from the collection system and brought to an adjacent biosafety cabinet for extraction and pooling into a viral suspension media.
7. All samples were sealed after collection and provided to lab staff for analysis after study completion.

Post Decontamination:

After each viral challenge test, the UV system inside the testing chamber was activated for 30 minutes. After 30 minutes of UV exposure, there was a 30-minute air purge through the air filtration system. Test equipment was cleaned at the end of each day with a 70% alcohol solution. Collection lines were soaked in a bleach bath mixture for 30 minutes then rinsed repeatedly with DI water. The nebulizer and vacuum collection pumps were decontaminated with hydrogen peroxide mixtures.

INNOVATIVE BIoANALYSIS

creating solutions | getting results

Preparation of The Pathogen

Viral Stock: SARS-CoV-2 USA-CA1/2020 (BEI NR-52382)

Test	Specifications	Results
Identification by Infectivity in Vero 6 cells	Cell Rounding and Detachment	Cell Rounding and Detachment
Next-Generation Sequencing (NGS) of the complete genome using Illumina® iSeq™ 100 Platform	≥ 98% identity with SARS-CoV 2, isolate USA-CA1/2020 GenBank: MN994467.1	99.9% identity with SARS-CoV 2, isolate USA-CA1/2020 GenBank: MN994467.1
Approx. 940 Nucleotides	≥ 98% identity with SARS-CoV 2, strain FDAARGOS_983 isolate USA-CA1/2020 GenBank: MT246667.1	100% identity with SARS-CoV 2, strain FDAARGOS_983 isolate USA-CA1/2020 GenBank: MT246667.1
Titer by TCID50 in Vero E6 Cells by cytopathic effect	Report Results	2.8 X 10 ⁵ TCID50 per mL in 5 days at 37°C and 5% CO ₂
Sterility (21-Day Incubation)		
Harpos HTYE Broth, aerobic	No Growth	No Growth
Trypticase Soy Broth, aerobic	No Growth	No Growth
Sabourad Broth, aerobic	No Growth	No Growth
Sheep Blood Agar, aerobic	No Growth	No Growth
Sheep Blood Agar, anaerobic	No Growth	No Growth
Thioglycollate Broth, anaerobic	No Growth	No Growth
DMEM with 10% FBS	No Growth	No Growth
Mycoplasma Contamination		
Agar and Broth Culture	None Detected	None Detected
DNA Detection by PCR of extracted test article nucleic acid	None Detected	None Detected

*The viral titer listed in the Certificate of Analysis is representative of the titer provided by BEI Resources. These viruses are grown on VeroE6 cells either in-house or at a partner lab to the concentrations listed within the experiment design.

INNOVATIVE BIOANALYSIS


creating solutions | getting results

Control Protocol

To accurately assess the Air Care O2-1 unit, a control group was conducted without the Air Care system running. Control samples were taken in the same manner and at the corresponding time points used for the challenge trial to serve as a comparative baseline to assess the viral reduction when the device was operating.

Study Results

Controls displayed a natural viability loss within the testing environment for 150 minutes. At 30 minutes against aerosolized SARS-CoV-2, the device decreased a 7.02×10^6 TCID₅₀/mL starting concentration to 8.46×10^5 TCID₅₀/mL. The concentration of collectible SARS-CoV-2 decreased over time with 9.60×10^3 TCID₅₀/mL at 60 minutes and reached below assay quantification levels after 90 minutes. The data showed that after 90 minutes of operation, the device achieved a 99.998% reduction.



**As it pertains to data represented herein, the value of 1.2E+02 indicates a titer that is lower than the specified limit of quantitation. The limit of quantitation for this assay is 1.2E+02.


***As it pertains to data represented; the percentage error equates to an average of $\pm 5\%$ of the final concentration.

INNOVATIVE BIOANALYSIS

creating solutions | getting results

VIRAL VOLUME REDUCTION OF SARS-COV-2		
	Control	Air Care O2-1
0 Minutes	7.02E+06	7.02E+06
30 Minutes	5.83E+06	8.46E+05
60 Minutes	4.96E+06	9.60E+03
90 Minutes	4.13E+06	1.20E+02
120 Minutes	3.22E+06	1.20E+02
150 Minutes	2.12E+06	1.20E+02

VIRAL LOG REDUCTION OF SARS-COV-2		
	Control	Air Care O2-1
0 Minutes	0.0000	0.0000
30 Minutes	0.0810	0.8379
60 Minutes	0.1505	2.7136
90 Minutes	0.2308	4.5364



Conclusion

The Tadiran Air Care O2-1 demonstrated a consistently higher reduction of aerosolized SARS-CoV-2 USA-CA1/2020 within a 1-m³ test box over 150 minutes compared to the natural loss rates observed during testing. The device achieved an 87.94% reduction after 30 minutes, 99.86% reduction after 60 minutes, and reached a 99.998% reduction after 90 minutes. As the test was designed to observe aerosol functions, it is unknown if any active pathogen remained on the surface areas inside the unit or the chamber walls. Furthermore, the study focused on the impact the Air Care O2-1 unit would have against an aerosolized pathogen in a specific volume of space. Therefore, when applied to a different size room, the results will scale and vary due to variables present, such as room size, air movement, and more. Every effort was made to simulate a replicable situation and address constraints with the experimental design and execution while taking the proper precautions when working with a BSL-3 pathogen.

INNOVATIVE BIOANALYSIS

creating solutions | getting results

DocuSigned by:

Dana Yee

7D5A69A0907947B...

12/13/2021

Dana Yee M.D

Date

Clinical Pathologist and Medical Director, Innovative Bioanalysis, Inc.

DocuSigned by:

Sam Kabbani

8B4B282DF4B34A3...

12/13/2021

Sam Kabbani, MS, BS, MT(ASCP), CLS

Date

Chief Scientific Officer, Innovative Bioanalysis, Inc.

DocuSigned by:

Albert Brockman

06DF5677A0D2400...

12/13/2021

Albert Brockman

Date

Chief Biosafety Officer, Innovative Bioanalysis, Inc.

DocuSigned by:

Kevin Noble

5DF2797BAA78421...

12/13/2021

Kevin Noble

Date

Laboratory Director, Innovative Bioanalysis, Inc.

[Disclaimer](#)

The Innovative Bioanalysis, Inc. ("Innovative Bioanalysis") laboratory is not certified or licensed by the United States Environmental Protection Agency and makes no equipment emissions claims pertaining to ozone or byproduct of any Tadiran Consumer & Technology Product, Ltd. device. Innovative Bioanalysis, Inc. makes no claims to the overall efficacy of any Air Care O2-1 device. The experiment results are solely applicable to the device used in the trial. The results are only representative of the experiment design described in this report. Innovative Bioanalysis, Inc. makes no claims as to the reproducibility of the experiment results given the possible variation of experiment results even with an identical test environment, viral strain, collection method, inoculation, nebulization, viral media, cell type, and culture procedure. Innovative Bioanalysis, Inc. makes no claims to third parties and takes no responsibility for any consequences arising out of the use of, or reliance on, the experiment results by third parties.